

**McGoff-Bethune, Inc.**  
5970 Unity Drive Suite A  
Norcross, GA 30071  
770-840-9811

# Certificate of Calibration

**McGoff-Bethune**  
5970 Unity Dr. Suite A  
Norcross, GA 30071

**Model of Standard:** Barber-Colman PA-39  
**Serial No.:** 0945-0002  
**Expiration Date:** 11/24/98

**Manufacturer:** Barber-Colman  
**Model:** 5642-02030-330-0-00  
**Input Range:** 0 to 2500 **Units:** Degrees F  
**Input Type:** TC-K **Serial No.:** X2927-2039  
**Tolerance:** +/- .25% of span (2 deg.) +/- 1LSD  
**Controller:**

<b>Input</b>	<b>As Found</b>	<b>As Left</b>	<b>Error</b>
100		100	
500	502	500	0
1000	1002	1000	0
1500	1500	1500	0
2000	1998	2000	0

**COMMENTS**

**Service Technician:** \_\_\_\_\_ **Date:** 4/6/98

This form when completed, signed and dated by the technician performing the certification is a verified document of indicated results. The above unit has been certified with an instrument that has been calibrated and tested using standards which are traceable to the National Institute of Standards and Technology, **per MIL-STD-45662A and conforms to 10CFR,Part 21.**